



Direct bilirubin measurements are used to monitor and diagnose **liver diseases**.

METHOD: DCA; ENDPOINT

FEATURES:

FOR THE QUANTITATIVE MEASUREMENT OF DIRECT BILIRUBIN

Genzyme Direct Bilirubin-SL Reagent

Genzyme's Direct Bilirubin-SL Reagent uses a spectrophotometric method that is reliable, convenient, and is intended for the measurement of direct bilirubin in serum.

- ✓ Two part stable liquid, ready to use reagent
- ✓ Applicable to multiple instrument platforms

BENEFITS:

- ✓ Easy to use, no additional preparation required
- ✓ Laboratory flexibility

ORDERING INFORMATION	CONFIGURATION	CATALOG NUMBER
<input type="radio"/> GENZYME DIRECT BILIRUBIN-SL	R1 1 X 100 mL R2 1 X 25 mL	247-10
<input type="radio"/> GENZYME DIRECT BILIRUBIN-SL	R1 3 X 100 mL R2 1 X 75 mL	247-30
<input type="radio"/> GENZYME DIRECT BILIRUBIN-SL	R1 1 X 1000 mL	247-50A
<input type="radio"/> GENZYME DIRECT BILIRUBIN-SL	R2 1 X 300 mL	247-50B
<input type="radio"/> GENZYME DC-CAL CALIBRATOR	5 X 3 mL	SE-035
<input type="radio"/> GENZYME DC-TROL LEVEL 1	10 X 5 mL	SM-052
<input type="radio"/> GENZYME DC-TROL LEVEL 2	10 X 5 mL	SM-056



THE AMERICAS
 31 New York Avenue
 Framingham, MA 01701-9322
 Phone: 800 332 1042
 Fax: 800 762 6311
 Email: usdiagcustomerservice@genzyme.com

INTERNATIONAL
 50 Gibson Drive
 Kings Hill, West Malling
 Kent, ME19 4AF, UK
 Phone: +44 (0) 1732 220022
 Fax: +44 (0) 1732 220024/5
 Email: ukdiagcustomerservice@genzyme.com

